



PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc. has assessed the Organization of:

R.M. Wester and Associates, Inc.

215 Indacom Dr., Saint Peters, MO 63376

*and hereby declares that the Organization is accredited in accordance with
the recognized International Standard:*

ISO/IEC 17025:2017

Whereby, technical competence has been confirmed for the associated scope supplement, in the fields of:

Ionizing Radiation and Radioactivity Calibration
(As detailed in the supplement)

Accreditation claims for conformity assessment activities shall only be made from the addresses referenced within this certificate and shall apply solely to those activities identified in the related scope. This Accreditation is granted subject to the Accreditation Body rules governing the Accreditation referred to above, and the Organization hereby commits to observing and complying with those rules in their entirety.

For PJLA:

Tracy Szerszen
President

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
755 W. Big Beaver, Suite 1325
Troy, Michigan 48084

Initial Accreditation Date:

September 22, 2026

Issue Date:

September 22, 2026

Expiration Date:

February 28, 2027


Accreditation No.:

138345

Certificate No.:

L26-845

*The validity of this certificate is maintained through ongoing assessments based
on a continuous accreditation cycle. The validity of this certificate should be
confirmed through the PJLA website: www.pjllabs.com*




Certificate of Accreditation: Supplement

R.M. Wester and Associates, Inc.

215 Indacom Dr., Saint Peters, MO 63376
Contact Name: Joseph D. Koch Phone: 636-928-9628

Accreditation is granted to the facility to perform the following conformity assessment activities:

FIELD OF CALIBRATION	MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	EXPANDED MEASUREMENT UNCERTAINTY ($\pm$) ¹	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED	FLEX CODE	LOCATION OF ACTIVITY
Ionizing Radiation and Radioactivity	Radiation Protection Survey Instruments (Cs-137 Air Kerma) (Dose Rate)	Up to 20 mSv/hr (Up to 2 000 mR/hr)	5.0 % of Reading	Cs-137 Sources JL Shepherd Series 10 Source Calibrator	CLab-001 "Calibration of Portable Handheld Instruments Using a Beam Port Source"	F0	F, O
Ionizing Radiation and Radioactivity	Radiation Protection Survey Instruments (Counts per Minute)	1 Cpm to 98 500 000 Cpm	2.0 % of Reading	Ludlum Model 500 Pulser or Equivalent	CLab-002 "Calibration of Portable Handheld Instruments Using a Pulse Generator"	F0	F, O
Ionizing Radiation and Radioactivity	Environmental Survey Instruments (Cs-137 Air Kerma) (Dose Rate)	Up to 20 mSv/hr (Up to 2 000 mR/hr)	5.0 % of Reading	Cs-137 Sources JL Shepherd Series 10 Source Calibrator	CLab-001 "Calibration of Portable Handheld Instruments Using a Beam Port Source"	F0	F, O
Ionizing Radiation and Radioactivity	Environmental Survey Instruments (Counts per Minute)	1 Cpm to 98 500 000 Cpm	2.0 % of Reading	Ludlum Model 500 Pulser or Equivalent	CLab-002 "Calibration of Portable Handheld Instruments Using a Pulse Generator"	F0	F, O
Ionizing Radiation and Radioactivity	Clinical Survey Instruments (Cs-137 Air Kerma) (Dose Rate)	Up to 20 mSv/hr (Up to 2 000 mR/hr)	5.0 % of Reading	Cs-137 Sources JL Shepherd Series 10 Source Calibrator	CLab-001 "Calibration of Portable Handheld Instruments Using a Beam Port Source"	F0	F
Ionizing Radiation and Radioactivity	Clinical Survey Instruments (Counts per Minute)	1 Cpm to 98 500 000 Cpm	2.0 % of Reading	Ludlum Model 500 Pulser or Equivalent	CLab-002 "Calibration of Portable Handheld Instruments Using a Pulse Generator"	F0	F
Ionizing Radiation and Radioactivity	Gate Monitor/ Portal Radiation Detection Instruments (Cs-137 Air Kerma) (Dose Rate)	Up to 20 mSv/hr (Up to 2 000 mR/hr)	5.0 % of Reading	Cs-137 Sources JL Shepherd Series 10 Source Calibrator	CLab-001 "Calibration of Portable Handheld Instruments Using a Beam Port Source"	F0	F, O
Ionizing Radiation and Radioactivity	Gate Monitor/ Portal Radiation Detection Instruments (Counts per Minute)	1 Cpm to 98 500 000 Cpm	2.0 % of Reading	Ludlum Model 500 Pulser or equivalent	CLab-002 "Calibration of Portable Handheld Instruments Using a Pulse Generator"	F0	F, O



Certificate of Accreditation: Supplement


R.M. Wester and Associates, Inc.

215 Indacom Dr., Saint Peters, MO 63376

Contact Name: Joseph D. Koch Phone: 636-928-9628

Accreditation is granted to the facility to perform the following conformity assessment activities:

FIELD OF CALIBRATION	MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	EXPANDED MEASUREMENT UNCERTAINTY ($\pm$) ¹	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED	FLEX CODE	LOCATION OF ACTIVITY
Ionizing Radiation and Radioactivity	Irradiation of Personnel Dosimeters Dose Equivalent (Cs-137 Air Kerma) (Dose Rate) (Electric)	Up to 20 mSv/hr (Up to 2 000 mR/hr)	4.9 % of Reading	Cs-137 Sources JL Shepherd Series 10 Source Calibrator	CLab-001 "Calibration of Portable Handheld Instruments Using a Beam Port Source"	F0	F
Ionizing Radiation and Radioactivity	Irradiation of Personnel Dosimeters Dose Equivalent (Cs-137 Air Kerma) (Dose) (Passive)	Up to 43.9 mGy (Up to 5 000 mR)	5.1 % of Reading	Cs-137 Sources JL Shepherd Series 10 Source Calibrator	CLab-001 "Calibration of Portable Handheld Instruments Using a Beam Port Source"	F0	F
Ionizing Radiation and Radioactivity	Thermoluminescent Dosimeter (TLD) Services (Cs-137 Air Kerma) (Dose Rate)	Up to 20 mSv/hr (Up to 2 000 mR/hr)	5.0 % of Reading	Cs-137 Sources JL Shepherd Series 10 Source Calibrator	CLab-001 "Calibration of Portable Handheld Instruments Using a Beam Port Source"	F0	F
Ionizing Radiation and Radioactivity	Thermoluminescent Dosimeter (TLD) Services (Cs-137 Air Kerma) (Dose)	Up to 43.9 mGy (Up to 5 000 mR)	5.0 % of Reading	Cs-137 Sources JL Shepherd Series 10 Source Calibrator	CLab-001 "Calibration of Portable Handheld Instruments Using a Beam Port Source"	F0	F
Ionizing Radiation and Radioactivity	Verification of Well Chambers for Nuclear Medicine (Cs-137)	Up to 3.7 MBq	4.2 % of Reading	Cs-137 Sources	CLab-DCWC01 "Verification of Dose Calibrators and Well Counters"	F0	F
Ionizing Radiation and Radioactivity	Verification of Well Chambers for Nuclear Medicine (Co-60)	Up to 7.4 MBq	4.8 % of Reading	CO-60 Sources	CLab-DCWC01 "Verification of Dose Calibrators and Well Counters"	F0	F
Ionizing Radiation and Radioactivity	Verification of Well Chambers for Nuclear Medicine (Ba-133)	Up to 500 MBq	3.9 % of Reading	Ba-133 Sources	CLab-DCWC01 "Verification of Dose Calibrators and Well Counters"	F0	F
Ionizing Radiation and Radioactivity	Portable Nuclear Density Gauge (Soil/Asphalt Density)	110.9 lb/ft ³ to 169.3 lb/ft ³ (1 777 kg/m ³ to 2 712 kg/m ³)	1.1 lb/ft ³	Calibration Blocks Humboldt Set Number 38	CLAB-013 "5 Block Calibration of Nuclear Density Gauges" ANSI N323AB-2013	F0	F



Certificate of Accreditation: Supplement

R.M. Wester and Associates, Inc.


215 Indacom Dr., Saint Peters, MO 63376
Contact Name: Joseph D. Koch Phone: 636-928-9628

Accreditation is granted to the facility to perform the following conformity assessment activities:

FIELD OF CALIBRATION	MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	EXPANDED MEASUREMENT UNCERTAINTY ($\pm$) ¹	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED	FLEX CODE	LOCATION OF ACTIVITY
Ionizing Radiation and Radioactivity	Portable Nuclear Density Gauge (Soil Moisture)	Up to 32.96 lb/ft ³ (Up to 528 kg/m ³)	1.3 lb/ft ³	Calibration Blocks Humboldt Set Number 38	CLAB-013 "5 Block Calibration of Nuclear Density Gauges" ANSI N323AB-2013	F0	F

1. The CMC (Calibration and Measurement Capability) is expressed in terms of measurement instrument/aspect being calibrated, range, expanded measurement uncertainty, equipment, and method/procedure. The expanded measurement uncertainty stated for calibrations included on this scope of accreditation represents the smallest measurement uncertainty attainable by the laboratory when performing a more or less routine calibration of a nearly ideal device under nearly ideal conditions. It is typically expressed at a confidence level of 95 % using a coverage factor k (usually equal to 2). The actual measurement uncertainty associated with a specific calibration performed by the laboratory will typically be larger than the measurement uncertainty included on this scope for the same calibration since capability and performance of the device being calibrated and the conditions related to the calibration may reasonably be expected to deviate from ideal to some degree.
2. The laboratory's range of calibration capability for all disciplines for which it is accredited is the interval from the smallest calibrated standard to the largest calibrated standard used in performing the calibration. The low end of this range must be an attainable value for which the laboratory has or has access to the standard referenced. Verification of an indicated value of zero in the absence of a standard is common practice in the procedure for many calibrations but by its definition it does not constitute calibration of zero capacity.
3. Location of activity:

Location Code	Location
F	Conformity assessment activity is performed at the CAB's fixed facility
O	Conformity assessment activity is performed onsite at the CAB's customer location
4. Measurement uncertainties obtained for calibrations performed at customer sites can be expected to be larger than the measurement uncertainties obtained at the laboratory's fixed location for similar calibrations. This is due to the effects of transportation of the standards and equipment and upon environmental conditions at the customer site which are typically not controlled as closely as at the laboratory's fixed location.



Certificate of Accreditation: Supplement

R.M. Wester and Associates, Inc.

215 Indacom Dr., Saint Peters, MO 63376

Contact Name: Joseph D. Koch Phone: 636-928-9628

Accreditation is granted to the facility to perform the following conformity assessment activities:

5. Flex Codes
F0: When no flexibility is identified. There are no changes to items calibrated, characteristics identified or versions of methods except for updating to the most recent version of a standard method after verification.
F1: The laboratory has the capability to introduce a new instrument, quantity, or gauge for an accredited calibration method.
F2: The laboratory has the capability to introduce the newest revision of an accredited authoritative standard method (with no modifications) identified on the scope
F3: The laboratory has the capability to introduce a new revision of an accredited non-standard method using the same technology or technique identified on the scope
F4: The laboratory has the capability to introduce a validated method that is equivalent to an accredited method (using the same Calibration Equipment or Reference Standards identified on the scope for the same parameter, component, or analyte identified on the line item of the scope.
- 6 The above scope of accreditation was created based on a former ILAC MRA Signatory's certificate policy. Based on the intent of the ILAC MRA, PJLA recognizes other scopes issued by other ILAC signatories. This scope will be modified based on PJLA's Policy following the next on-site assessment.

